



Australian Government

Department of Health
Therapeutic Goods Administration

Australian Register of Therapeutic Goods Certificate

Issued to

FOTEK

for approval to supply

FOTEK - Electrosurgical system, general-purpose

ARTG Identifier	348110
ARTG Start Date	12/11/2020
Product Category	Medical Device Included Class IIb
GMDN	44776
GMDN Term	Electrosurgical system, general-purpose
Intended Purpose	Electrosurgical systems "FOTEK" is an assembly of devices intended to generate, and apply radio-frequency (RF) alternating current to soft tissues for cutting and coagulation during an endoscopic or open surgical procedure. The assembly includes a mains electricity-powered (AC-powered) RF generator with monitoring functions/controls, connection cables, a handpiece, and electrodes to apply the energy to the surgical site. Tissue resistance to the electrical current creates the heat as the current travels through the body between electrodes. "FOTEK" generators and accessories are compatible with electrosurgical accessories and generators of all modern manufactures.

Manufacturer Details	Address	Certificate number(s)
Fotek OOO	Malisheva Street 145 A liter A , Ekaterinburg , 620049 Russian Federation	DV-2020-MC-30153-1

ARTG Standard Conditions

The above Medical Device Included Class IIb has been entered on the Register subject to the following conditions:

- The inclusion of the kind of device in the ARTG is subject to compliance with all conditions placed or imposed on the ARTG entry. Refer Part 4-5, Division 2 (Conditions) of the Therapeutic Goods Act 1989 and Part 5, Division 5.2 (Conditions) of the Therapeutic Goods (Medical Devices) Regulations 2002 for relevant information.
- Breaching conditions of the inclusion related to the device of the kind may lead to suspension or cancellation of the ARTG entry; may be a criminal offence; and civil penalties may apply.

Products Covered by This Entry

1. Electrosurgical system, general-purpose

Product Specific Conditions

No specific conditions have been recorded against this entry.

Therapeutic Goods Administration
PO Box 100, Woden ACT 2606 Australia
Phone: 1800 020 653
Email: info@tga.gov.au

ARTG Identifier: 348110
ARTG Start Date: 12/11/2020